

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004147  
10:22 04/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:47  
PRESSURIZING 2:50  
STERILIZING 3:56  
136.4 C 220kPa 3:56  
136.4 C 225kPa 4:26  
136.4 C 224kPa 4:56  
136.4 C 224kPa 5:26  
136.4 C 224kPa 5:56  
136.4 C 224kPa 6:26  
136.0 C 221kPa 6:56  
136.4 C 223kPa 7:26  
136.4 C 223kPa 7:26  
VENTING 7:26  
AIR DRYING 8:19  
CYCLE COMPLETE 8:20

-----  
Digital Signature #  
A79A93DC8D1ACE8C  
-----