
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001040
10:15 19/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:55
STERILIZING 5:25
136.4 C 226kPa 5:25
136.3 C 224kPa 5:55
136.4 C 225kPa 6:25
136.3 C 222kPa 6:55
135.9 C 222kPa 7:25
136.4 C 222kPa 7:55
136.2 C 223kPa 8:25
136.1 C 221kPa 8:55
136.1 C 221kPa 8:55
VENTING 8:55
AIR DRYING 9:51
CYCLE COMPLETE 10:33

Digital Signature #
1CF759929115FF1C
