
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004152
14:25 04/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:40
STERILIZING 3:41
136.4 C 221kPa 3:41
135.9 C 221kPa 4:11
136.3 C 224kPa 4:41
136.4 C 224kPa 5:11
136.3 C 222kPa 5:41
136.4 C 224kPa 6:11
136.3 C 220kPa 6:41
136.4 C 225kPa 7:11
136.4 C 225kPa 7:11
VENTING 7:11
AIR DRYING 8:00
CYCLE COMPLETE 8:02

Digital Signature #
F8C068796C064E8F
