

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001321  
15:56 26/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:43  
PRESSURIZING 3:44  
STERILIZING 4:58  
136.0 C 220kPa 4:58  
136.2 C 224kPa 5:28  
136.4 C 225kPa 5:58  
136.3 C 224kPa 6:28  
136.4 C 225kPa 6:58  
136.2 C 223kPa 7:28  
136.3 C 225kPa 7:58  
136.4 C 224kPa 8:28  
136.4 C 224kPa 8:28  
VENTING 8:28  
AIR DRYING 9:23  
CYCLE COMPLETE 9:49

-----  
Digital Signature #  
AF15AB10A299D591  
-----