
Midmark M11 - v1.0.0
Total Cycles: 1266

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2021 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.2 F 0.0
2:00 187.8 F 3.4
4:00 193.5 F 8.1
6:00 211.1 F 9.1
8:00 226.7 F 8.1
10:00 239.7 F 13.5
12:00 240.5 F 11.3
14:00 239.3 F 10.1
16:00 254.0 F 17.4
18:00 266.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.6 F 28.8
2:00 271.8 F 28.5
3:00 272.7 F 28.9
4:00 271.6 F 28.2

Min 271.6 F 28.2
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.4 F 0.0
5:00 209.2 F 0.0
10:00 209.6 F 0.0
15:00 208.6 F 0.0
20:00 204.0 F 0.0
25:00 199.6 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

10/14/2021 11:42 AM
ACCEPTED: X
REJECTED:
Operator: U02