

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3299

Name: RITTER2  
Sterilizer ID: V2313228  
01/15/2024 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.3 F 0.0  
2:00 156.7 F 2.2  
4:00 177.9 F 5.7  
6:00 197.9 F 10.4  
8:00 221.8 F 5.4  
10:00 225.0 F 9.9  
12:00 232.4 F 8.3  
14:00 241.1 F 12.8  
16:00 242.2 F 11.7  
18:00 254.6 F 18.4  
20:00 266.4 F 25.3  
22:00 273.0 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.7 F 29.7  
2:00 271.8 F 29.7  
3:00 272.6 F 29.6  
4:00 271.8 F 29.7

Min 271.5 F 29.2  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 168.7 F 0.0  
5:00 250.0 F 0.0  
10:00 194.4 F 0.0  
15:00 190.9 F 0.0  
20:00 186.7 F 0.0  
25:00 183.1 F 0.0  
30:00 180.8 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:23  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

01/15/2024 01:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_