
Midmark M11 - v1.0.5
Total Cycles: 1557

Name: RITTER2
Sterilizer ID: V2313228
10/04/2022 07:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.3 F 0.0
2:00 80.7 F 0.0
4:00 110.9 F 0.0
6:00 147.2 F 0.0
8:00 161.3 F 0.0
10:00 190.7 F 0.3
12:00 207.7 F 2.8
14:00 218.0 F 6.2
16:00 228.4 F 9.9
18:00 235.4 F 10.5
20:00 238.9 F 11.1
22:00 236.8 F 9.2
24:00 239.7 F 9.9
26:00 249.2 F 14.9
28:00 261.2 F 21.7
30:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.9
2:00 272.2 F 29.1
3:00 272.0 F 28.6
4:00 272.5 F 29.3

Min 271.5 F 28.4
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 211.6 F 0.0
10:00 233.2 F 0.0
15:00 227.8 F 0.0
20:00 225.0 F 0.0
25:00 224.1 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:29
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

10/04/2022 08:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____