
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002347
13:15 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:05
STERILIZING 4:10
136.4 C 225kPa 4:10
136.2 C 224kPa 4:40
136.3 C 225kPa 5:10
136.4 C 226kPa 5:40
136.4 C 226kPa 6:10
136.4 C 226kPa 6:40
136.4 C 225kPa 7:10
136.3 C 226kPa 7:40
136.3 C 226kPa 7:40
VENTING 7:40
AIR DRYING 8:33
CYCLE COMPLETE 31:30

Digital Signature #
907B2168E75A9624
