

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000412  
13:10 03/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:15  
STERILIZING 4:17  
136.4 C 224kPa 4:17  
136.4 C 223kPa 4:47  
136.2 C 223kPa 5:17  
136.3 C 224kPa 5:47  
136.3 C 224kPa 6:17  
136.4 C 223kPa 6:47  
136.3 C 224kPa 7:17  
136.2 C 222kPa 7:47  
136.2 C 222kPa 7:47  
VENTING 7:47  
AIR DRYING 8:38  
CYCLE COMPLETE 9:24

-----  
Digital Signature #  
9D6789A58FD27655  
-----