
Midmark M11 - v1.0.5
Total Cycles: 2292

Name: RITTER2
Sterilizer ID: V2313228
04/24/2023 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.9 F 0.0
2:00 186.4 F 4.4
4:00 206.8 F 10.2
6:00 221.6 F 5.9
8:00 235.7 F 12.1
10:00 237.2 F 10.5
12:00 237.7 F 9.8
14:00 252.7 F 17.2
16:00 265.1 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.8 F 29.6
2:00 271.5 F 29.0
3:00 272.7 F 29.7
4:00 272.0 F 29.1

Min 271.4 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 227.7 F 0.0
10:00 233.1 F 0.0
15:00 226.1 F 0.0
20:00 223.1 F 0.0
25:00 223.4 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 17:44
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:11
CYCLE COMPLETE

04/24/2023 02:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____