
Midmark M11 - v1.0.5
Total Cycles: 1953

Name: RITTER2
Sterilizer ID: V2313228
01/19/2023 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 184.3 F 4.4
4:00 196.6 F 9.4
6:00 215.5 F 6.9
8:00 230.7 F 9.9
10:00 233.8 F 8.5
12:00 243.7 F 12.4
14:00 244.9 F 13.1
16:00 257.4 F 20.0
18:00 268.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.1 F 29.5
2:00 271.3 F 28.8
3:00 271.9 F 29.7
4:00 272.0 F 29.3

Min 271.1 F 28.7
Max 272.2 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 236.3 F 0.0
10:00 232.2 F 0.0
15:00 228.4 F 0.0
20:00 228.6 F 0.0
25:00 227.6 F 0.0
30:00 229.6 F 0.0

Drying Complete

FILLING: 1:58
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

01/19/2023 04:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____