
Midmark M11 - v1.0.5
Total Cycles: 3310

Name: RITTER2
Sterilizer ID: V2313228
01/19/2024 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 173.1 F 2.9
4:00 189.8 F 7.6
6:00 209.0 F 13.2
8:00 223.0 F 8.7
10:00 234.5 F 9.9
12:00 241.7 F 12.7
14:00 242.8 F 12.0
16:00 256.1 F 19.0
18:00 268.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 273.1 F 29.6
2:00 272.8 F 30.1
3:00 271.9 F 29.4
4:00 273.0 F 30.0

Min 271.5 F 29.3
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 238.8 F 0.0
10:00 200.7 F 0.0
15:00 190.0 F 0.0
20:00 186.6 F 0.0
25:00 184.4 F 0.0
30:00 182.6 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

01/19/2024 02:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____