
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005148
9:39 27/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:45
STERILIZING 3:54
136.4 C 223kPa 3:54
136.4 C 222kPa 4:24
136.4 C 224kPa 4:54
136.2 C 219kPa 5:24
136.2 C 222kPa 5:54
136.4 C 223kPa 6:24
136.4 C 223kPa 6:54
136.0 C 220kPa 7:24
136.0 C 220kPa 7:24
VENTING 7:24
AIR DRYING 8:12
CYCLE COMPLETE 10:46

Digital Signature #
7230523A02B4E6F2
