

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3509

Name: RITTER 1  
Sterilizer ID: V2178198  
07/11/2023 07:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 73.7 F 0.0  
4:00 81.7 F 0.0  
6:00 102.6 F 0.0  
8:00 134.9 F 1.0  
10:00 170.5 F 2.9  
12:00 184.5 F 5.9  
14:00 198.2 F 9.1  
16:00 214.2 F 12.5  
18:00 219.2 F 5.8  
20:00 230.4 F 9.8  
22:00 240.5 F 11.3  
24:00 238.9 F 10.6  
26:00 234.4 F 7.8  
28:00 248.0 F 14.3  
30:00 258.9 F 20.2  
32:00 268.6 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.7 F 29.1  
2:00 271.5 F 28.1  
3:00 272.3 F 29.1  
4:00 271.9 F 28.5  
  
Min 271.2 F 28.1  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.4 F 0.0  
5:00 203.8 F 0.0  
10:00 201.0 F 0.0  
15:00 203.7 F 0.0  
20:00 206.0 F 0.0  
25:00 204.9 F 0.0  
30:00 207.2 F 0.0  
  
Drying Complete

FILLING: 0:47  
HEATING: 33:14  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:11:12  
CYCLE COMPLETE

07/11/2023 08:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_