
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000496
11:09 30/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:50
STERILIZING 3:56
136.0 C 229kPa 3:56
136.1 C 230kPa 4:26
136.4 C 231kPa 4:56
136.4 C 232kPa 5:26
136.3 C 231kPa 5:56
136.4 C 232kPa 6:26
136.4 C 233kPa 6:56
136.4 C 233kPa 7:26
136.4 C 233kPa 7:26
VENTING 7:26
AIR DRYING 8:15
CYCLE COMPLETE 16:10

Digital Signature #
4DB2B073D462FC67
