

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4138

Name: RITTER2  
Sterilizer ID: V2313228  
11/19/2024 04:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.2 F 0.0  
2:00 141.0 F 0.1  
4:00 167.3 F 2.0  
6:00 195.3 F 6.4  
8:00 209.0 F 11.3  
10:00 224.7 F 6.3  
12:00 232.8 F 11.3  
14:00 234.2 F 8.6  
16:00 240.2 F 10.6  
18:00 246.7 F 13.8  
20:00 258.9 F 20.5  
22:00 269.9 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 272.3 F 29.3  
2:00 272.6 F 29.6  
3:00 272.6 F 29.9  
4:00 271.7 F 29.3

Min 271.7 F 29.1  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
5:00 216.3 F 0.0  
10:00 189.6 F 0.0  
15:00 184.5 F 0.0  
20:00 179.7 F 0.0  
25:00 177.6 F 0.0  
30:00 177.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 23:39  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:02:03  
CYCLE COMPLETE

11/19/2024 05:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_