
Midmark M11 - v1.0.0
Total Cycles: 4593

Name: RITTER 1
Sterilizer ID: V2178198
05/29/2024 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 160.9 F 2.3
4:00 167.3 F 5.6
6:00 189.6 F 9.4
8:00 206.5 F 13.0
10:00 218.4 F 6.0
12:00 230.5 F 10.5
14:00 230.8 F 7.6
16:00 241.1 F 12.0
18:00 238.8 F 10.0
20:00 251.0 F 16.0
22:00 261.8 F 22.1
24:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.1
2:00 271.8 F 28.5
3:00 271.5 F 28.7
4:00 272.5 F 29.1

Min 271.2 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 231.5 F 0.0
10:00 190.4 F 0.0
15:00 185.1 F 0.0
20:00 182.9 F 0.0
25:00 183.1 F 0.0
30:00 189.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:10
CYCLE COMPLETE

05/29/2024 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____