

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001011  
13:59 16/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:55  
PRESSURIZING 1:58  
STERILIZING 2:59  
136.4 C 231kPa 2:59  
136.3 C 233kPa 3:29  
136.4 C 233kPa 3:59  
136.4 C 232kPa 4:29  
136.5 C 233kPa 4:59  
136.2 C 231kPa 5:29  
136.4 C 233kPa 5:59  
136.3 C 233kPa 6:29  
136.3 C 233kPa 6:29  
VENTING 6:29  
AIR DRYING 7:21  
CYCLE COMPLETE 7:22

-----  
Digital Signature #  
535310BBD64D3170  
-----