
Midmark M11 - v1.0.0
Total Cycles: 5489

Name: RITTER 1
Sterilizer ID: V2178198
04/08/2025 04:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.6 F 0.0
2:00 104.0 F 1.1
4:00 124.2 F 3.7
6:00 152.7 F 6.8
8:00 192.6 F 10.2
10:00 210.9 F 2.8
12:00 211.9 F 5.7
14:00 224.3 F 9.6
16:00 239.8 F 10.8
18:00 238.0 F 10.3
20:00 237.0 F 8.6
22:00 246.4 F 13.7
24:00 257.4 F 19.5
26:00 267.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.1 F 29.0
2:00 271.2 F 28.3
3:00 272.1 F 29.2
4:00 271.6 F 28.5

Min 271.1 F 28.3
Max 272.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 193.6 F 0.0
10:00 200.8 F 0.0
15:00 193.0 F 0.0
20:00 190.4 F 0.0
25:00 193.8 F 0.0
30:00 205.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:25
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:05:58
CYCLE COMPLETE

04/08/2025 05:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____