

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1952

Name: RITTER2  
Sterilizer ID: V2313228  
01/19/2023 02:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.4 F 0.0  
2:00 154.1 F 2.8  
4:00 170.2 F 6.3  
6:00 199.2 F 11.0  
8:00 212.8 F 5.2  
10:00 224.7 F 10.5  
12:00 229.5 F 7.1  
14:00 242.5 F 13.5  
16:00 242.8 F 12.4  
18:00 255.2 F 19.0  
20:00 266.7 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.6 F 30.2  
2:00 272.0 F 29.5  
3:00 271.6 F 29.6  
4:00 272.6 F 30.1

Min 271.4 F 29.2  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 217.8 F 0.0  
10:00 229.9 F 0.0  
15:00 235.1 F 0.0  
20:00 226.4 F 0.0  
25:00 218.9 F 0.0  
30:00 217.4 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 21:32  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:01  
CYCLE COMPLETE

01/19/2023 03:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_