
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000508
11:31 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:39
STERILIZING 4:48
136.5 C 224kPa 4:48
136.3 C 222kPa 5:18
136.4 C 225kPa 5:48
136.1 C 222kPa 6:18
136.4 C 224kPa 6:48
136.4 C 225kPa 7:18
136.0 C 222kPa 7:48
136.1 C 223kPa 8:18
136.1 C 223kPa 8:18
VENTING 8:18
AIR DRYING 9:11
CYCLE COMPLETE 9:18

Digital Signature #
9277A29D02BE1FCD
