
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001635
16:02 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:22
STERILIZING 3:25
136.0 C 220kPa 3:25
136.4 C 226kPa 3:55
136.4 C 225kPa 4:25
136.2 C 224kPa 4:55
136.4 C 223kPa 5:25
136.2 C 224kPa 5:55
136.1 C 223kPa 6:25
136.3 C 225kPa 6:55
136.3 C 225kPa 6:55
VENTING 6:55
AIR DRYING 7:48
CYCLE COMPLETE 7:52

Digital Signature #
3CC26C18FEE20B3E
