

-----  
Midmark M11 - v1.0.5  
Total Cycles: 899

Name: RITTER2  
Sterilizer ID: V2313228  
03/24/2022 08:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.3 F 0.0  
2:00 117.6 F 0.0  
4:00 124.8 F 0.0  
6:00 169.7 F 0.7  
8:00 191.3 F 4.5  
10:00 215.1 F 8.7  
12:00 226.2 F 13.5  
14:00 230.7 F 9.5  
16:00 235.1 F 9.5  
18:00 243.9 F 13.4  
20:00 243.8 F 12.6  
22:00 256.4 F 19.3  
24:00 267.5 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.7 F 29.7  
2:00 271.7 F 29.0  
3:00 271.6 F 29.5  
4:00 272.6 F 29.6

Min 271.1 F 28.8  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 218.5 F 0.0  
10:00 228.9 F 0.0  
15:00 226.8 F 0.0  
20:00 226.6 F 0.0  
25:00 216.6 F 0.0  
30:00 202.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:24  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:03:31  
CYCLE COMPLETE

03/24/2022 09:47 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02