

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4564

Name: RITTER 1  
Sterilizer ID: V2178198  
05/16/2024 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.7 F 0.0  
2:00 159.7 F 3.5  
4:00 181.2 F 7.8  
6:00 205.3 F 12.3  
8:00 217.3 F 6.0  
10:00 231.0 F 11.0  
12:00 230.1 F 7.3  
14:00 243.5 F 13.3  
16:00 242.4 F 11.7  
18:00 255.4 F 18.3  
20:00 266.9 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.6 F 29.1  
2:00 271.5 F 28.4  
3:00 272.2 F 29.1  
4:00 272.4 F 28.9  
  
Min 271.2 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 208.9 F 0.0  
10:00 204.9 F 0.0  
15:00 230.3 F 0.0  
20:00 231.7 F 0.0  
25:00 230.5 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 21:30  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:00:11  
CYCLE COMPLETE

05/16/2024 03:29 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_