
Midmark M11 - v1.0.5
Total Cycles: 1209

Name: RITTER2
Sterilizer ID: V2313228
06/22/2022 07:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 79.8 F 0.0
4:00 89.8 F 0.0
6:00 112.4 F 0.4
8:00 146.4 F 3.1
10:00 173.0 F 6.3
12:00 195.6 F 9.9
14:00 212.8 F 6.7
16:00 218.4 F 7.2
18:00 229.6 F 11.2
20:00 229.3 F 7.2
22:00 241.9 F 12.8
24:00 240.6 F 11.2
26:00 252.7 F 17.3
28:00 263.3 F 23.6
30:00 272.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.6 F 29.8
2:00 271.8 F 29.0
3:00 271.6 F 29.5
4:00 272.4 F 29.7

Min 271.1 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 201.7 F 0.0
10:00 218.1 F 0.0
15:00 208.7 F 0.0
20:00 211.8 F 0.0
25:00 210.6 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:18
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

06/22/2022 08:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____