
Midmark M11 - v1.0.5
Total Cycles: 4113

Name: RITTER2
Sterilizer ID: V2313228
11/13/2024 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.5 F 0.0
2:00 145.6 F 0.0
4:00 154.1 F 1.8
6:00 185.6 F 4.9
8:00 201.0 F 8.6
10:00 215.0 F 12.4
12:00 221.5 F 7.2
14:00 231.9 F 11.5
16:00 232.8 F 8.3
18:00 243.1 F 12.6
20:00 243.4 F 12.3
22:00 254.6 F 18.2
24:00 264.7 F 24.5
26:00 270.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 271.9 F 29.7
2:00 271.8 F 29.8
3:00 271.8 F 29.8
4:00 271.7 F 29.7

Min 271.7 F 29.5
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 219.9 F 0.0
10:00 226.8 F 0.0
15:00 221.4 F 0.0
20:00 220.9 F 0.0
25:00 226.1 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:41
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:27
CYCLE COMPLETE

11/13/2024 01:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____