
Midmark M11 - v1.0.5
Total Cycles: 2249

Name: RITTER2
Sterilizer ID: V2313228
04/13/2023 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.4 F 0.0
2:00 185.4 F 3.7
4:00 197.5 F 8.6
6:00 214.4 F 10.6
8:00 229.0 F 9.2
10:00 235.5 F 10.1
12:00 242.7 F 12.9
14:00 242.6 F 12.1
16:00 255.4 F 18.7
18:00 266.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.8 F 29.6
2:00 271.3 F 28.9
3:00 272.4 F 29.8
4:00 271.8 F 29.2

Min 271.2 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 214.2 F 0.0
10:00 200.8 F 0.0
15:00 200.6 F 0.0
20:00 199.2 F 0.0
25:00 197.7 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

04/13/2023 03:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____