
Midmark M11 - v1.0.0
Total Cycles: 1452

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2021 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 75.9 F 0.0
4:00 95.7 F 0.0
6:00 133.8 F 0.1
8:00 164.5 F 0.6
10:00 184.2 F 2.4
12:00 192.8 F 4.5
14:00 207.6 F 7.0
16:00 217.7 F 10.2
18:00 223.6 F 9.7
20:00 229.6 F 8.4
22:00 239.7 F 13.0
24:00 238.9 F 10.2
26:00 237.3 F 8.4
28:00 249.1 F 14.6
30:00 260.0 F 20.8
32:00 270.3 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.1
3:00 272.5 F 28.8
4:00 271.6 F 28.0

Min 271.3 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.1 F 0.0
5:00 198.5 F 0.0
10:00 208.4 F 0.0
15:00 206.6 F 0.0
20:00 204.4 F 0.0
25:00 209.9 F 0.0
30:00 215.6 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 32:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:11:04
CYCLE COMPLETE

12/14/2021 08:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____