
Midmark M11 - v1.0.0
Total Cycles: 2767

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2023 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.2 F 0.0
2:00 186.4 F 0.0
4:00 212.9 F 4.1
6:00 222.1 F 9.3
8:00 232.2 F 9.1
10:00 238.5 F 11.0
12:00 236.2 F 8.6
14:00 240.1 F 9.8
16:00 249.4 F 14.5
18:00 261.8 F 21.5
20:00 272.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.5 F 28.5
2:00 271.3 F 27.8
3:00 272.4 F 28.7
4:00 272.0 F 28.2

Min 271.2 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 237.3 F 0.0
10:00 233.4 F 0.0
15:00 230.9 F 0.0
20:00 228.1 F 0.0
25:00 231.5 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 20:21
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

01/11/2023 02:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____