

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5466

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2025 04:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.4 F 0.0  
2:00 166.9 F 3.8  
4:00 182.5 F 8.6  
6:00 200.3 F 9.4  
8:00 219.5 F 6.8  
10:00 232.8 F 11.8  
12:00 233.0 F 8.5  
14:00 239.9 F 10.3  
16:00 245.3 F 13.1  
18:00 258.2 F 20.0  
20:00 269.7 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.2 F 28.7  
2:00 272.8 F 29.1  
3:00 272.4 F 29.1  
4:00 272.0 F 28.5  
  
Min 271.7 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 194.6 F 0.0  
10:00 208.2 F 0.0  
15:00 213.6 F 0.0  
20:00 214.0 F 0.0  
25:00 214.2 F 0.0  
30:00 211.3 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:23  
CYCLE COMPLETE

03/31/2025 05:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_