
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000064
11:46 03/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:52
136.4 C 225kPa 4:52
136.4 C 224kPa 5:22
136.4 C 226kPa 5:52
136.4 C 222kPa 6:22
136.3 C 222kPa 6:52
136.1 C 224kPa 7:22
136.4 C 226kPa 7:52
136.4 C 226kPa 8:22
136.4 C 226kPa 8:22
VENTING 8:22
AIR DRYING 9:16
CYCLE COMPLETE 15:36

Digital Signature #
9A472C199C493A2E
