

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001997  
8:50 10/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:21  
PRESSURIZING 3:22  
STERILIZING 4:29  
136.4 C 225kPa 4:29  
136.2 C 224kPa 4:59  
136.3 C 224kPa 5:29  
136.2 C 223kPa 5:59  
136.1 C 222kPa 6:29  
136.4 C 225kPa 6:59  
136.3 C 225kPa 7:29  
136.3 C 222kPa 7:59  
136.3 C 222kPa 7:59  
VENTING 7:59  
AIR DRYING 8:52  
CYCLE COMPLETE 9:50  
-----

Digital Signature #  
A2F9F38E861DA464  
-----