

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4594

Name: RITTER 1  
Sterilizer ID: V2178198  
05/29/2024 12:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 94.1 F 0.0  
2:00 94.9 F 0.6  
4:00 107.4 F 2.4  
6:00 124.3 F 4.8  
8:00 152.8 F 7.7  
10:00 175.4 F 10.9  
12:00 201.7 F 5.3  
14:00 212.5 F 6.0  
16:00 222.9 F 9.5  
18:00 232.5 F 13.4  
20:00 233.4 F 9.3  
22:00 242.7 F 11.3  
24:00 241.8 F 11.7  
26:00 252.6 F 17.0  
28:00 262.4 F 22.7  
30:00 271.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.7  
1:00 272.4 F 29.4  
2:00 272.3 F 29.3  
3:00 272.6 F 29.7  
4:00 271.6 F 29.1

Min 271.0 F 28.6  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 215.9 F 0.0  
10:00 226.5 F 0.0  
15:00 228.1 F 0.0  
20:00 230.7 F 0.0  
25:00 223.5 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:38  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:09:14  
CYCLE COMPLETE

05/29/2024 01:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_