
Midmark M11 - v1.0.5
Total Cycles: 3276

Name: RITTER2
Sterilizer ID: V2313228
01/04/2024 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.2 F 0.0
2:00 161.2 F 1.9
4:00 180.4 F 5.3
6:00 200.0 F 9.6
8:00 216.9 F 9.5
10:00 225.2 F 7.9
12:00 238.2 F 12.9
14:00 239.0 F 10.7
16:00 237.1 F 9.5
18:00 251.3 F 16.2
20:00 262.8 F 22.9
22:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.1 F 29.5
2:00 272.0 F 28.9
3:00 271.4 F 29.1
4:00 273.2 F 30.0

Min 270.9 F 28.6
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 238.3 F 0.0
10:00 206.1 F 0.0
15:00 188.0 F 0.0
20:00 184.1 F 0.0
25:00 182.0 F 0.0
30:00 179.5 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:05
CYCLE COMPLETE

01/04/2024 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____