
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001399
16:30 04/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:53
STERILIZING 5:01
136.0 C 220kPa 5:01
136.4 C 225kPa 5:31
136.4 C 225kPa 6:01
136.3 C 224kPa 6:31
136.4 C 222kPa 7:01
136.3 C 224kPa 7:31
136.1 C 222kPa 8:01
136.4 C 223kPa 8:31
136.4 C 223kPa 8:31
VENTING 8:31
AIR DRYING 9:28
CYCLE COMPLETE 13:32

Digital Signature #
6BB95308DC702247
