
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006466
8:54 16/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:57
STERILIZING 2:58
136.4 C 223kPa 2:58
136.1 C 218kPa 3:28
136.3 C 224kPa 3:58
136.4 C 223kPa 4:28
136.4 C 224kPa 4:58
136.4 C 223kPa 5:28
136.3 C 220kPa 5:58
136.4 C 221kPa 6:28
136.4 C 221kPa 6:28
VENTING 6:28
AIR DRYING 7:22
CYCLE COMPLETE 9:23

Digital Signature #
A1D4EC410EA3F234
