
Midmark M11 - v1.0.5
Total Cycles: 2070

Name: RITTER2
Sterilizer ID: V2313228
02/21/2023 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 194.8 F 2.9
4:00 207.3 F 8.8
6:00 225.7 F 8.9
8:00 235.6 F 11.7
10:00 237.2 F 10.4
12:00 237.5 F 9.9
14:00 252.9 F 17.4
16:00 265.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.2 F 29.5
2:00 271.2 F 29.1
3:00 272.3 F 29.6
4:00 271.3 F 28.9

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 228.2 F 0.0
10:00 243.4 F 0.0
15:00 241.6 F 0.0
20:00 224.9 F 0.0
25:00 225.1 F 0.0
30:00 232.9 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 17:36
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:07
CYCLE COMPLETE

02/21/2023 11:40 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____