
Midmark M11 - v1.0.0
Total Cycles: 4382

Name: RITTER 1
Sterilizer ID: V2178198
03/15/2024 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 157.6 F 2.1
4:00 164.5 F 5.1
6:00 191.8 F 8.6
8:00 213.0 F 13.2
10:00 216.6 F 6.1
12:00 227.5 F 10.3
14:00 233.7 F 8.1
16:00 240.5 F 11.6
18:00 238.3 F 9.5
20:00 250.5 F 15.4
22:00 260.9 F 21.2
24:00 270.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 271.6 F 28.6
2:00 271.9 F 28.7
3:00 272.1 F 28.9
4:00 272.1 F 29.1

Min 271.2 F 28.6
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 206.6 F 0.0
10:00 197.5 F 0.0
15:00 193.4 F 0.0
20:00 192.2 F 0.0
25:00 191.0 F 0.0
30:00 188.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:33
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:22
CYCLE COMPLETE

03/15/2024 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____