

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005706  
7:34 08/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:33  
STERILIZING 3:37  
136.4 C 220kPa 3:37  
136.3 C 220kPa 4:07  
136.2 C 219kPa 4:37  
136.4 C 222kPa 5:07  
136.4 C 222kPa 5:37  
136.5 C 223kPa 6:07  
136.2 C 220kPa 6:37  
136.1 C 221kPa 7:07  
136.1 C 221kPa 7:07  
VENTING 7:07  
AIR DRYING 7:56  
CYCLE COMPLETE 8:10

-----  
Digital Signature #  
C9BDB17DDF52829B  
-----