
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001160
12:13 08/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:42
STERILIZING 4:49
136.0 C 222kPa 4:49
136.2 C 225kPa 5:19
136.3 C 224kPa 5:49
136.2 C 223kPa 6:19
136.2 C 223kPa 6:49
136.2 C 223kPa 7:19
136.4 C 226kPa 7:49
136.1 C 222kPa 8:19
136.1 C 222kPa 8:19
VENTING 8:19
AIR DRYING 9:15
CYCLE COMPLETE 24:15

Digital Signature #
F7B1DE989E5C28FD
