
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000332
12:35 18/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:08
STERILIZING 4:31
136.0 C 228kPa 4:31
136.1 C 231kPa 5:01
136.3 C 233kPa 5:31
136.4 C 232kPa 6:01
136.3 C 232kPa 6:31
136.1 C 231kPa 7:01
136.4 C 233kPa 7:31
136.4 C 228kPa 8:01
136.4 C 228kPa 8:01
VENTING 8:01
AIR DRYING 8:56
CYCLE COMPLETE 23:56

Digital Signature #
85D1E7552A5C96AA
