

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1180

Name: RITTER2  
Sterilizer ID: V2313228  
06/09/2022 09:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
2:00 172.8 F 2.8  
4:00 177.2 F 7.0  
6:00 200.5 F 12.6  
8:00 225.4 F 8.1  
10:00 238.3 F 12.2  
12:00 238.1 F 10.9  
14:00 236.2 F 9.4  
16:00 250.7 F 16.2  
18:00 262.5 F 23.1  
20:00 273.0 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.4  
1:00 272.9 F 30.0  
2:00 271.4 F 29.4  
3:00 273.0 F 30.0  
4:00 271.7 F 29.4

Min 271.4 F 29.3  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 232.6 F 0.0  
10:00 224.5 F 0.0  
15:00 232.4 F 0.0  
20:00 228.4 F 0.0  
25:00 224.9 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 21:13  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:37  
CYCLE COMPLETE

06/09/2022 10:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_