
Midmark M11 - v1.0.5
Total Cycles: 3790

Name: RITTER2
Sterilizer ID: V2313228
07/16/2024 03:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.7 F 0.0
2:00 129.0 F 1.3
4:00 145.2 F 4.0
6:00 170.0 F 7.6
8:00 192.1 F 12.0
10:00 217.2 F 5.9
12:00 227.8 F 10.9
14:00 227.6 F 7.1
16:00 240.7 F 13.1
18:00 241.2 F 11.8
20:00 253.5 F 18.1
22:00 265.1 F 24.8
24:00 271.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.9
1:00 271.9 F 30.3
2:00 272.2 F 30.0
3:00 271.8 F 29.8
4:00 271.4 F 29.9

Min 271.4 F 29.8
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 188.7 F 0.0
10:00 185.6 F 0.0
15:00 184.7 F 0.0
20:00 183.6 F 0.0
25:00 183.5 F 0.0
30:00 185.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:47
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:35
CYCLE COMPLETE

07/16/2024 04:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____