
Midmark M11 - v1.0.5
Total Cycles: 3447

Name: RITTER2
Sterilizer ID: V2313228
03/08/2024 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.4 F 0.0
2:00 133.1 F 1.1
4:00 159.1 F 4.3
6:00 184.4 F 8.7
8:00 206.4 F 13.5
10:00 219.1 F 6.7
12:00 230.7 F 11.5
14:00 232.4 F 8.3
16:00 240.8 F 11.3
18:00 245.2 F 13.3
20:00 258.0 F 20.1
22:00 269.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.2 F 29.6
2:00 272.3 F 29.1
3:00 271.7 F 29.2
4:00 272.4 F 29.5

Min 271.2 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.9 F 0.0
5:00 197.2 F 0.0
10:00 200.6 F 0.0
15:00 200.0 F 0.0
20:00 197.6 F 0.0
25:00 195.5 F 0.0
30:00 193.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

03/08/2024 03:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____