
Midmark M11 - v1.0.0
Total Cycles: 2221

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2022 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 186.2 F 3.7
4:00 186.3 F 8.1
6:00 211.2 F 12.6
8:00 223.4 F 6.4
10:00 229.9 F 11.2
12:00 231.9 F 7.3
14:00 243.8 F 13.1
16:00 241.9 F 11.2
18:00 253.9 F 17.4
20:00 264.9 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.9
2:00 271.3 F 28.2
3:00 272.4 F 29.1
4:00 272.0 F 28.5

Min 271.2 F 28.2
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 223.5 F 0.0
10:00 206.5 F 0.0
15:00 206.0 F 0.0
20:00 225.1 F 0.0
25:00 228.8 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

08/15/2022 11:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____