
Midmark M11 - v1.0.0
Total Cycles: 1576

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 04:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.6 F 0.0
2:00 181.2 F 3.6
4:00 186.6 F 8.4
6:00 211.0 F 9.8
8:00 224.5 F 8.0
10:00 237.8 F 13.4
12:00 239.2 F 10.8
14:00 237.7 F 9.1
16:00 252.1 F 16.2
18:00 264.4 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 28.8
2:00 271.6 F 28.1
3:00 272.9 F 29.0
4:00 272.1 F 28.3

Min 271.5 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 222.5 F 0.0
10:00 237.8 F 0.0
15:00 228.5 F 0.0
20:00 225.7 F 0.0
25:00 227.4 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

01/31/2022 05:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____