

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3108

Name: RITTER2  
Sterilizer ID: V2313228  
11/09/2023 03:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.0 F 0.0  
2:00 177.1 F 3.5  
4:00 198.9 F 8.7  
6:00 219.4 F 9.5  
8:00 228.4 F 10.0  
10:00 234.1 F 8.3  
12:00 242.2 F 11.9  
14:00 246.1 F 13.6  
16:00 259.2 F 20.7  
18:00 270.7 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.4 F 30.0  
2:00 271.5 F 29.4  
3:00 272.6 F 29.5  
4:00 272.5 F 30.0  
  
Min 270.9 F 28.7  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 236.8 F 0.0  
10:00 207.9 F 0.0  
15:00 191.1 F 0.0  
20:00 186.8 F 0.0  
25:00 184.5 F 0.0  
30:00 183.2 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 18:37  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 00:57:56  
CYCLE COMPLETE

11/09/2023 04:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_