
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000873
15:46 17/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:56
STERILIZING 4:11
136.0 C 227kPa 4:11
136.4 C 233kPa 4:41
136.4 C 228kPa 5:11
136.4 C 233kPa 5:41
136.2 C 231kPa 6:11
136.2 C 231kPa 6:41
136.2 C 232kPa 7:11
136.4 C 231kPa 7:41
136.4 C 231kPa 7:41
VENTING 7:41
AIR DRYING 8:34
CYCLE COMPLETE 21:16

Digital Signature #
643854BCD1593B40
