
Midmark M11 - v1.0.0
Total Cycles: 921

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2021 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 143.1 F 1.2
4:00 154.0 F 3.6
6:00 204.1 F 7.1
8:00 211.7 F 12.0
10:00 222.4 F 5.3
12:00 232.5 F 9.9
14:00 238.1 F 9.0
16:00 241.7 F 12.0
18:00 239.1 F 10.2
20:00 252.0 F 16.5
22:00 263.1 F 22.9
24:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.2
2:00 271.8 F 28.5
3:00 272.0 F 29.0
4:00 272.4 F 29.0

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 218.4 F 0.0
10:00 202.8 F 0.0
15:00 214.4 F 0.0
20:00 231.9 F 0.0
25:00 231.9 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

06/21/2021 12:26 PM
ACCEPTED: X
REJECTED:
Operator: U01