
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000773
15:08 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:09
CASSETTE REMOVED
NOT STERILE
END TIME 0:09

Digital Signature #
3B600EA2762D1944
